

Comparison of working length determination with radiographs and two electronic apex locators

J. P. Vieyra¹, J. Acosta² & J. M. Mondaca²

¹School of Dentistry, Universidad Autónoma de Baja California, Tijuana, Baja California, México; and ²Private Practice in Endodontics, 710E San Ysidro Blvd., 1513 San Ysidro, California 92173, USA

Abstract

Vieyra JP, Acosta J, Mondaca JM. Comparison of working length determination with radiographs and two electronic apex locators. *International Endodontic Journal*, **43**, 16–20, 2010.

Aim To evaluate the accuracy of the Root ZX and Elements-Diagnostic electronic apex locators when compared with radiographs for locating the canal terminus or minor foramen.

Methodology The canal terminus of 482 canals in 160 maxillary and mandibular teeth was located *in vivo* with both locators and radiographically. After extraction, the actual location of the minor foramen was determined visually and with magnification. A paired samples *t*-test, chi-square test and a repeated measure ANOVA at the 0.05 level of significance were used to determine differences between the groups.

Results The Root ZX located the minor foramen correctly 68% of the time in anterior and premolar teeth, and 58% of the time in molar teeth. The Elements-Diagnostic located the minor foramen

correctly 58% of the time in anterior and premolar teeth and 49% of the time in molar teeth. Radiographs located the minor foramen correctly 20% of the time in anterior and premolar teeth and 11% of the time in molar teeth. There was no statistically significant difference between the two locators, but there was a significant difference between them and radiographs. For all teeth, the measurements made by the apex locators were within ± 0.5 mm of the minor foramen 100% of the time, whereas for the radiographs, the measurements were within this range only 15% of the time. This difference was significant ($P = 0.05$).

Conclusion Measuring the location of the minor foramen using the two apex locators was more accurate than radiographs and would reduce the risk of instrumenting and filling beyond the apical foramen.

Keywords: apical constriction, electronic apex locator, elements-diagnostic, Root ZX, working length determination.

Received 20 November 2008; accepted 8 July 2009

Introduction

Root canal preparation and filling should not extend beyond the tooth root nor leave uninstrumented areas inside the root canal. Anatomically, the apical constriction (AC), also called the minor apical diameter or minor diameter (Kuttler 1955), is a logical location for working length (WL), as it often coincides with the narrowest diameter of the root canal (AAE 2003).

However, locating the AC clinically is problematic. Dummer *et al.* (1984) concluded that it is impossible to locate the minor foramen clinically with certainty because of its position and topography. The cemento-dentinal junction (CDJ) has also been suggested as the location for WL, because it represents the transition between pulpal and periodontal tissue (Grove 1931). The location of the CDJ is widely accepted as being 0.50–0.75 mm coronal to the apical foramen (Ricucci & Langeland 1998) but, as with the AC, the exact location of the CDJ is impossible to identify clinically. In general, the CDJ is considered to be co-located with the minor foramen (Stein *et al.* 1990); however, this is not always the case (Dummer *et al.* 1984).

Correspondence: Dr Jorge Paredes Vieyra, PMB#1513, 710E, San Ysidro Blvd., Suite "A", San Ysidro, CA 92173, USA (Tel.: +1 619 946 0459; fax: +1 664 687 2207; e-mail: jorgitoparedesvieyra@hotmail.com).

Working length is defined as 'the distance from a coronal reference point to the point at which canal preparation and filling should terminate' (American Association of Endodontists (AAE) 2003). Radiographic determination of WL has limitations such as distortion, shortening and elongation, interpretation variability and lack of three-dimensional representation. Even when a paralleling technique is used, elongation of images has been found to be approximately 5% (Van de Voorde & Bjondahl 1969).

A WL 1 mm short of the radiographic apex may result in over or under instrumentation because of the variability in distance between the terminus of the root canal (minor foramen) and the radiographic apex (Gutiérrez & Aguayo 1995). Thus, this often used 'rule' is not predictable or reliable.

Custer (1918) was the first to determine WL electronically. Suzuki (1942) investigated the electrical resistance properties of oral tissues and developed the first electronic apex locator (EAL). The device was resistance-based and measured the resistance between two electrodes to determine the location of an instrument in the canal. Later devices were impedance-based (Nekoofar *et al.* 2006) and used multiple frequencies. More recently, resistance- and capacitance-based devices emerged that measure resistance and capacitance, directly and independently.

The Root ZX (J. Morita Corp., Tokyo, Japan) uses the 'ratio method' to locate the minor foramen (Kobayashi & Suda 1994) by the simultaneous measurement of impedance using two frequencies. The Root ZX claims to work in the presence of electrolytes and nonelectrolytes and requires no calibration (Kobayashi 1995).

The Elements-Diagnostic (Sybron Endo, Sybron Dental, Orange, CA, USA) uses multiple frequencies, in an attempt to eliminate the influence of canal conditions.

In addition to improving WL accuracy (Nekoofar *et al.* 2006), EALs address concerns about radiation, as they have the potential to reduce the number of radiographs taken during root canal treatment (Pagavino *et al.* 1998).

The purpose of this study was to evaluate *in vivo* the accuracy and predictability of two EALs for determining WL as compared with radiographs.

Materials and methods

One hundred and sixty teeth (482 canals) with fully formed apices and without apical resorption were used (Table 1). All teeth gave positive responses to hot and cold tests and were extracted for periodontal or

prosthodontic reasons. Ethical approval for the study and an informed consent to participate was signed by the patients.

After local anaesthesia, rubber dam isolation and access cavity preparation were performed, the canals were flared coronally with size 1 and 2 Orifice Shapers (Dentsply Tulsa Dental, Tulsa, OK, USA) using 3% sodium hypochlorite (NaOCl) for irrigation. The final rinse was aspirated, but no attempt was made to dry the canals.

The AC of each tooth was located with two EALs and radiographically.

The minor foramen was located with the Root ZX by advancing a size 15 stainless steel K-file in the canal, until the locator indicated that the minor foramen had been reached, according to the manufacturer's instructions (J. Morita Corp. 2004). The LCD showed a flashing bar between APEX and 1 and a flashing tooth. The silicone stop on the file was positioned at the reference point. The file was removed from the canal and the length was measured to the nearest 0.01 mm with a digital caliper. This was the insertion length.

The AC was located with the Elements-Diagnostic EAL by advancing the same size 15 K-file in the canal, until the locator indicated that the minor foramen had been reached, as per the manufacturer's instructions (Sybron Endo 2003). The stop was positioned at the reference point and the insertion length measured. The sequence of testing alternated between the two locators.

The minor foramen was located radiographically by advancing the size 15 K-file, until its tip was 1.0 mm from the radiographic apex (determined from a pre-treatment parallel technique radiograph). A radiograph was exposed and if the file tip was seen not to be 1.0 mm from the radiographic apex, the file was repositioned and another radiograph taken to ensure that it was. The distance from the stop to the tip was the insertion length. The file was then re-inserted to the insertion length (1 mm from the radiographic

Table 1 Distribution of 160 teeth (482 canals)

Tooth (<i>n</i>)	Number of canals	
	Maxillary	Mandibular
Central incisor (10)	7	3
Lateral incisor (8)	6	2
Canine (5)	3	2
Premolar (17)	11	6
Molar (120)	225	217
Total (160)	252	230

apex) and cemented in place with Fuji II LC dual-cure glass ionomer cement (GC Corp, Tokyo, Japan). The file handle was sectioned with a high-speed bur and the tooth was extracted without disturbing the file, placed in 6% NaOCl for 15 min to clean the root surface and stored in a 0.2% thymol solution. All of the clinical procedures were conducted by the principal investigator.

After the tooth was removed from the solution and with the file still in place, the apical 5 mm of the root was ground parallel to the long axis of the canal with a fine diamond bur and abrasive discs. When the file

became visible, additional dentine was removed under 20 × magnification (OPMI Pico microscope; Carl Zeiss, Munich, Germany) until the file tip, the canal terminus, and the foramen were in focus. A digital photograph was taken and stored in Adobe Photoshop 5.5 (Adobe Systems Inc., San Jose, CA, USA) and the distance of the file tip to the minor foramen was measured. This distance was recorded as being: -1.0 mm from the minor foramen; -0.5 mm from the minor foramen; at the minor foramen; +0.5 mm from the minor foramen or +1.0 mm from the minor foramen. A minus symbol (-) indicated a file short of the minor foramen; a plus symbol (+) indicated it was long.

Once the actual length to the minor foramen was measured visually, the distance from the minor foramen determined by the two EALs was also completed (-1.0 mm from the minor foramen; -0.5 mm from the minor foramen, etc.), by comparing their insertion lengths to the actual length (distance to the AC) (Tables 2–4).

The measurements obtained by the two EALs and radiographs relative to the actual location of the minor foramen were compared using a paired samples *t*-test, chi-square test and a repeated measure. ANOVA evaluation was conducted at the 0.05 level of significance.

Table 2 Distance of file tip from minor foramen determined by Root ZX, Elements and radiograph (anterior)

Distance from minor foramen (mm)	Root ZX <i>n</i> = 23 (%)	Elements <i>n</i> = 23 (%)	Radiograph <i>n</i> = 23 (%)
-1.0	-	-	-
-0.5	-	-	-
MF	17 (73.9)	15 (65.2)	5 (21.7)
+0.5	6 (26.08)	8 (34.7)	10 (43.47)
+1.0			8 (34.78)

MF, minor foramen.

(+) and (-) values indicate file tip beyond (+) or short (-) of the AC.

Table 3 Distance of file tip from minor foramen determined by Root ZX, Elements and radiograph (premolars)

Distance from minor foramen (mm)	Root ZX <i>n</i> = 17 (%)	Elements <i>n</i> = 17 (%)	Radiograph <i>n</i> = 17 (%)
-1.0	-	-	-
-0.5	-	-	-
MF	9 (52.94)	7 (41.17)	6 (35.29)
+0.5	8 (47.05)	10 (58.82)	5 (29.41)
+1.0			6 (35.29)

MF, minor foramen.

(+) and (-) values indicate file tip beyond (+) or short (-) of the AC.

Table 4 Distance of file tip from minor foramen determined by Root ZX, Elements and radiograph (molars)

Distance from minor foramen (mm)	Root ZX (<i>n</i> = 444)						Elements (<i>n</i> = 423)						Radiograph (<i>n</i> = 414)					
	Canal						Canal						Canal					
	MB	ML	D	DB	DL	Pa	MB	ML	D	DB	DL	Pa	MB	ML	D	DB	DL	Pa
-1.0	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
-0.5	2	3	-	-	-	-	6	5	18	-	-	8	1	-	2	-	-	1
MF	65	61	58	19	19	38	66	59	54	18	19	2	11	8	12	6	5	5
+0.5	53	45	27	16	16	22	48	41	15	22	21	21	61	41	47	17	17	28
+1.0	-	-	-	-	-	-	-	-	-	-	-	-	43	34	26	11	12	26

MF, minor foramen.

(+) and (-) values indicate file tip beyond (+) or short (-) of the AC.

Results

For anterior teeth, the Root ZX, Elements and radiographs located the minor foramen 74%, 65% and 22% of the time, respectively. For premolar teeth, the Root ZX, Elements and radiographs located the minor foramen 53%, 41% and 35% of the time, respectively. For molar teeth, the Root ZX, Elements and radiographs located the minor foramen 58%, 49% and 11% of the time, respectively. There was no statistically significant difference between the two EALs, but there was a difference when the EALs and radiographs were compared (Tables 2–4).

For anterior, premolar and molar teeth, none of the measurements were 1.0 mm short of the minor foramen. For anterior and premolar teeth, none of the measurements were 0.5 mm short of the minor foramen, but for molar teeth 1%, 8% and 1% of the measurements using the Root ZX, Elements and radiographs, respectively, were short.

For anterior teeth, the Root ZX, Elements and radiographs were 0.5 mm long of the minor foramen a 26%, 35% and 39% roots, respectively. For premolar teeth, the Root ZX, Elements and radiographs were 0.5 mm long of the minor foramen 47%, 59% and 29% roots, respectively, and for molar teeth it was 41%, 42% and 48%, respectively.

No EAL measurements were 1.0 mm long of the minor foramen for anterior, premolar and molar teeth, but for radiographs it was 35% for anterior teeth, 35% for premolar teeth and 37% for molar teeth. There was no statistically significant difference between the two EALs, but there was a significant difference ($P = 0.05$) when the EALs and radiographs were compared.

Discussion

The use of electronic devices to determine WL has gained in popularity. When using them, an important consideration is being aware of the possible sources of error such as metallic restorations, salivary contamination, dehydration, etc. However, as shown in this and other studies, the accuracy of EALs is superior to radiographs (Van de Voorde & Bjondahl 1969, Pratten & McDonald 1996, Venturi & Breschi 2007).

One of the reasons why a radiographically determined WL lacks accuracy is that it is based on the radiographic apex rather than the canal terminus – the minor foramen. WL is obtained with a radiograph by positioning the tip of a file a certain distance (usually 1.0 mm) from the radiographic apex. However, WL should be based on the location of the minor foramen rather than the apex, because the foramen frequently is not at the apex (Wrbas et al. 2007). In this study, radiographs correctly located the minor foramen 15% of the time, whereas for the Root ZX and Elements it was 63% and 53% of the time, respectively. Both EALs were within ± 0.5 mm from the minor foramen 100% of the time, whereas radiographs were within ± 0.5 mm of 63% of cases. An *in vivo* study by Shabahang et al. (1996) reported that the Root ZX was within 0.5 mm from the minor foramen 96% of the time, a value

similar to the present findings. In general, this study also agrees with others (Usun et al. 2007, 2008) that EALs are more accurate than radiographs and greatly reduce the risk of instrumenting and filling short or beyond the canal terminus.

As the minor foramen varies in location and anatomy (sharply defined, parallel, or missing) (Nekoofar et al. 2006), caution should be used to avoid overestimating WL. According to Gutiérrez & Aguayo (1995), over-instrumentation of the root canal must be a common and unnoticed occurrence. An instrument passing through a necrotic pulp and through the foramen most likely carries bacteria and toxins into the apical tissues (Siqueira et al. 2002, Siqueira & Barnett 2004). An indication by an EAL of reaching the minor foramen or foramen is very helpful in avoiding mishaps. Indeed this study showed that WL obtained with radiographs was 1.0 mm long of the AC 37% of the time, but 0% for the two EALs. This high incidence of error is clinically important, because a WL 1.0 mm long would result in canals being instrumented beyond the foramen.

Conclusion

Under clinical conditions, the EALs identified the minor foramen with high degree of accuracy. EALs were more accurate compared with radiographs with the potential to greatly reduce the risk of instrumenting and filling beyond the apical foramen.

Acknowledgements

We thank Dr E. Steve Senia and Dr Michael Hülsmann for their valuable assistance in reviewing this manuscript.

References

- American Association of Endodontists (AAE) (2003) *Glossary of Endodontic Terms*, 7th edn. Chicago, IL: American Association of Endodontists.
- Custer LE (1918) Exact method of locating the apical foramen. *Journal of the National Dental Association* **5**, 815–9.
- Dummer PMH, McGinn JH, Rees DG (1984) The position and topography of the apical canal constriction and apical foramen. *International Endodontic Journal* **17**, 192–8.
- Grove CJ (1931) The value of the dentinocemental junction in pulp canal surgery. *Journal of Dental Research* **11**, 466–8.
- Gutiérrez JH, Aguayo P (1995) Apical foraminal openings in Human teeth. Number and location. *Oral Surgery, Oral*

- Medicine, Oral Pathology, Oral Radiology, and Endodontics* **79**, 769–77.
- J. Morita Corp. (2004) *Fully Automatic Root Canal Measuring Device. Root ZX Operation*. Tokyo: J. Morita Corp.
- Kobayashi C (1995) Electronic canal length measurement. *Oral Surgery, Oral Medicine, Oral Pathology, Oral Radiology, and Endodontics* **79**, 177–9.
- Kobayashi C, Suda H (1994) New electronic canal measuring device based on the ratio method. *Journal of Endodontics* **20**, 111–4.
- Kuttler Y (1955) Microscopic investigation of root apices. *Journal of the American Dental Association* **50**, 544–52.
- Nekoofar MN, Ghandi MM, Hayes SJ, Dummer PMH (2006) The fundamental operating principles of electronic root canal length measurement devices. *International Endodontic Journal* **39**, 595–609.
- Pagavino G, Pace R, Baccetti T (1998) An SEM study of in vivo accuracy of the Root ZX electronic apex locator. *Journal of Endodontics* **24**, 438–41.
- Pratten DH, McDonald NJ (1996) Comparison of radiographic and electronic working lengths. *Journal of Endodontics* **22**, 173–6.
- Ricucci D, Langeland K (1998) Apical limit of root canal instrumentation and obturation, part 2: a histological study. *International Endodontic Journal* **31**, 394–409.
- Shabahang S, Goon WWY, Gluskin AH (1996) An in vivo evaluation of Root ZX electronic apex locator. *Journal of Endodontics* **22**, 616–8.
- Siqueira JF, Barnett F (2004) Interappointment pain: mechanisms, diagnosis, and treatment. *Endodontic Topics* **7**, 93–109.
- Siqueira JF Jr, Rôças IN, Favieri A, et al. (2002) Incidence of postoperative pain after intracanal procedures based on an antimicrobial strategy. *Journal of Endodontics* **6**, 457–60.
- Stein TJ, Corcoran JF, Zillich RM (1990) The influence of the major and minor foramen diameters on apical electronic probe measurements. *Journal of Endodontics* **16**, 520–2.
- Suzuki K (1942) Experimental study on iontophoresis. *Journal of the Japanese Stomatology* **16**, 411.
- Sybron Endo (2003) *Elements Diagnostic: Instruction Guidelines*. Glendora, CA: Sybron Endo, pp. 1–3.
- Usun O, Topuz O, Tinaz AC, Sadik B (2007) Accuracy of the apex locating function of TCM Endo V in simulated conditions: a comparison study. *Oral Surgery, Oral Medicine, Oral Pathology, Oral Radiology, and Endodontics* **103**, e73–6.
- Usun O, Topuz O, Tinaz C, Nekoofar MH, Dummer PMH (2008) Accuracy of two root canal length measurement devices integrated into rotary endodontic motors when removing gutta-percha from root-filled teeth. *International Endodontic Journal* **41**, 725–32.
- Van de Voorde HE, Bjondahl AM (1969) Estimating endodontic “working length” with paralleling radiographs. *Oral Surgery, Oral Medicine, and Oral Pathology* **27**, 106–10.
- Venturi M, Breschi L (2007) A comparison between two electronic apex locators: an *ex vivo* investigation. *International Endodontic Journal* **40**, 362–73.
- Wrbas KT, Ziegler AA, Altenburger MJ, Schirrmeister JF (2007) In vivo comparison of working length determination with two electronic apex locators. *International Endodontic Journal* **40**, 133–8.